



LANTERN LARYNGEAL ELECTRODE

Complete Cord Contact
For effective Recurrent
Laryngeal Nerve monitoring

- ✓ Conforms to vocal cord position every time
- ✓ Gentle self-regulating pressure to vocal cords
- ✓ Offers excellent discrimination between vocal cords
- ✓ Suitable for use with multi-channel nerve monitors
- ✓ Available to fit 4mm to 9mm endotracheal tubes
- ✓ Electrode tail does not obscure anaesthetist's view

Neurosign's patented Lantern Laryngeal Electrode is a radical new design for recurrent laryngeal & vagus nerve monitoring. The revolutionary lantern design of the electrode ensures contact with the vocal cords in all patients, even when the endotracheal tube position is suboptimal.

Designed in collaboration with Dr. Jack Kartush



Contact your Neurosign®
representative for further
information.

www.neurosignsurgical.com
info@neurosignsurgical.com





THE CONTRAST IN DESIGN BETWEEN THE LANTERN LARYNGEAL ELECTRODE & A STANDARD LARYNGEAL ELECTRODE




Figure A: The Neurosign® Lantern Laryngeal Electrode, with a dynamic electrode that optimizes electrode contact

Figure B: A standard laryngeal electrode with flush mounted, static electrodes that result in reduced vocal cord contact during surgery

LANTERN LARYNGEAL ELECTRODE SIZING AND ORDERING INFORMATION:

- 4200-00**
Lantern Laryngeal Electrode 8/9mm, box of 10, sterile, single use, including a reusable connection cable
- 4201-00**
Lantern Laryngeal Electrode 6/7mm, box of 10, sterile, single use, including a reusable connection cable
- 4202-00**
Lantern Laryngeal Electrode 4/5mm, box of 1, sterile, single use, including a reusable connection cable
- 4203-00**
Reusable connection cable (N-Series), box of 1, non-sterile
- 4203-NM**
Reusable connection cable (Neurosign V4), box of 1, non-sterile

- ✓ Suitable for use with any standard or reinforced endotracheal tube
- ✓ Allows monitoring of Recurrent Laryngeal Nerve even when trachea is distorted
- ✓ Excellent discrimination between vocal cords
- ✓ Can be used with multi-channel nerve monitors
- ✓ Time saving setup with integrated ground electrode
- ✓ Low movement artefact promotes clearer signals

NMK1703-03

© 2017, The Magstim Company Limited. Magstim® and Neurosign® are registered trademarks of The Magstim Company Limited. All specifications are subject to change. The information contained within this document is accurate at the time of publication.

 Technomed
Amerikalaan 71, 6199 AE Maastricht-Airport, The Netherlands

 **NEUROSIGN®**
www.neurosignsurgical.com